

# Technical Data Sheet

Doc No – Ab/TDS/LH-57-I

Rev No : 03

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**Pipette Tips, Bulk PP**

**Valid for Cat. No. P10115, P10137, P10129, P10130, P10101, P10105, P10138, P10102, P10106**

### 1. Description/Specification

**1.1 Description:** Disposable PP Micropipette Tips, DNase, RNase and Pyrogen Free.

#### 1.2 Design:

1.2.1 Open Cylindrical Upper and conical Apex with orifice for liquid dispensing

#### 1.3 Performance Specifications:

Code	Capacity	Colour	No. per Case
P10115	0.2-10µl (Graduated)	Natural	1000
P10137	0.2-10µl Extended (Graduated)	Natural	1000
P10129	2-200µl (Bevelled Graduated)	Yellow	1000
P10130	2-200µl (Bevelled Graduated)	Natural	1000
P10101	2-200µl (Ultra Point Tip Orifice)	Yellow	1000
P10105	2-200µl (Ultra Point Tip Orifice)	Natural	1000
P10138	2-300µl (Graduated)	Natural	1000
P10102	100 - 1000µl	Blue	500
P10106	100 - 1000µl	Natural	500

1.3.2 Fitting to pipettes: Universal Fit (For detail refer the compatibility chart)

#### 1.4 Material:



Shiv Ganga Industrial Estate,  
Plot no. 1 Lakeshwari, Bhagwanpur,  
Roorkee - 247667 Uttarakhand, India  
**An ISO 9001:2008 Company**  
A Division of Abdos

**Regd. Office :**  
13A, Government Place (East)  
Kolkata – 700069, India  
**CIN :** U85100WB2008PTC123903

**P :** +91 (0) 1332 235277  
**E :** info@abdoslaware.com  
**W :** www.abdoslaware.com

# Technical Data Sheet

1.4.1 Polypropylene (PP), Hydrophobic

1.4.2 This material meets the FDA criteria Regulation 21 CFR 177.1520. olefin polymers.

1.4.3 Free from heavy metals according to ELV Direction 2000/53/EC and its following amendments.

1.4.4 This material is used for all devices for diagnostic (Class-1 medical devices) according to Directive 93/42 requirements.

## 1.5 Color:

**Tip color-** Natural, Yellow and Blue

## 1.6 Sterilization:

Sterile catalog number from this lot have been pre-sterilized and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation.

## 1.7 Quality control:

Testing of various characteristics in accordance with the valid Specification.

## 1.8 Other information:

Reusable, only box can be autoclavable up to 10 autoclave cycle

Tips are of single used only.

## 1.9 Basic Features:

**1.9.1. Made up of :** High Purity Virgin Polypropylene (PP)

**1.9.2. Temperature range :** -80<sup>0</sup>C to + 121<sup>0</sup>C for 20 minutes

**1.9.3. Autoclavability:** Yes, Autoclavable at 15 psi and 121<sup>0</sup>C for 20 minutes

**1.9.4. Chemical resistance:** see website.


**1.9.5. Retention:** Best in its class. (with Acidic & Basic dyes)

## Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121<sup>0</sup>C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50<sup>0</sup>C before removing product.



# Technical Data Sheet

  
Quality Assurance Incharge  
Abdos Labtech Pvt. Ltd.



Shiv Ganga Industrial Estate,  
Plot no. 1 Lakeshwari, Bhagwanpur,  
Roorkee - 247667 Uttrakhand, India  
**An ISO 9001:2008 Company**  
A Division of Abdos

**Regd. Office :**  
13A, Government Place (East)  
Kolkata - 700069, India  
**CIN :** U85100WB2008PTC123903

**P :** +91 (0) 1332 235277  
**E :** info@abdoslabware.com  
**W :** www.abdoslabware.com