

Technical Data Sheet

Doc No – Ab/TDS/CT-07

Rev No : 05

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Self-Standing Centrifuge Tubes, PP/HDPE

Valid For Cat. No. P10423, P10424, P10450, P10432, P10449

1. Description/Specification

1.1 Description:

Disposable PP/HDPE centrifuge tube, DNase, RNase and Pyrogen Free

1.2 Design:

1.2.1 Tube: Cylindrical Upper and conical Bottom with self-stand (skirted)

1.2.2 Tube Cap: Separate Threaded HDPE Cap

1.3 Performance Specifications:

1.3.1 Volume: **P10423** (50ml Non Sterile), **P10424** (50ml Sterile), **P10450** (50ml Sterile Red), **P10432** (50ml Blue Bluk), **P10449** (50ml Red)

1.4 Material:

1.4.1 **Tube:** Polypropylene (PP), Hydrophobic **Cap**-HDPE

1.4.2 Polypropylene (PP), Hydrophobic, confirm US FDA 21CFR

1.4.3 Complies with requirement accordance with article of regulation (EC) No. 1907/2006 and its amendments.

1.4.4 Complies with the requirement of European Pharmacopoeia, 7th edition (2011) & its supplements 7.5 (07/2012) monographs 3.2.2 Plastic container and closures for Pharmaceutical use

1.4.5 The product has passed the USP testing including class VI tests & has been assigned the FDA Drug master file DMF 12123

1.4.6 Material has Successful Passed the biological tests according to ISO 10993 external Communication devices for indirect blood contact for Prolonged period.

1.5 Color:

Tube color- Translucent (Natural)

Cap color- Blue/Red



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttrakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com

Technical Data Sheet

1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.6.2 White Patch for easy writing on the tube

1.7 Sterilization:

Sterile catalog number from this lot have been pre-sterilized and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation.

1.8 Quality control:

DNase, RNase, and Pyrogen Free

Testing of various characteristics in accordance with the valid Specification

1.9 Other information

Recommended for single use only

Pack Size: 500pieces/case

2. Basic Features

2.1 Temperature range: Tube:-80⁰C to +121⁰C

2.2 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

2.3 Centrifugation: 3500 RCF (Rated with Double Distilled water at 10⁰C for 30 minutes)

2.4 Leak Test at 635 mm Hg: No leakage

2.5 Chemical resistance: See website.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.





Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttarakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com

Technical Data Sheet


Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttarakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata - 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com